

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4567

Name: RITTER 1  
Sterilizer ID: V2178198  
05/21/2024 09:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.4 F 0.0  
2:00 174.6 F 2.2  
4:00 175.8 F 5.7  
6:00 204.5 F 9.4  
8:00 219.9 F 7.6  
10:00 218.9 F 7.9  
12:00 231.1 F 12.7  
14:00 233.0 F 9.3  
16:00 239.4 F 9.9  
18:00 245.1 F 13.2  
20:00 256.6 F 19.3  
22:00 267.2 F 25.7  
24:00 272.9 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.2  
1:00 272.9 F 29.7  
2:00 271.7 F 29.1  
3:00 272.8 F 29.8  
4:00 271.9 F 29.1

Min 271.7 F 29.1  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.6 F 0.0  
5:00 224.3 F 0.0  
10:00 229.1 F 0.0  
15:00 228.4 F 0.0  
20:00 229.7 F 0.0  
25:00 231.8 F 0.0  
30:00 230.9 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 24:23  
STERILIZING: 4:00  
VENTING: 3:41  
DRYING: 30:00  
TOTAL CYCLE: 01:03:10  
CYCLE COMPLETE

05/21/2024 10:41 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_