
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002508
7:29 25/10/2021

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:37
STERILIZING 3:40
136.4 C 223kPa 3:40
135.9 C 221kPa 4:10
136.2 C 221kPa 4:40
136.1 C 223kPa 5:10
136.3 C 222kPa 5:40
136.1 C 219kPa 6:10
136.2 C 217kPa 6:40
136.4 C 223kPa 7:10
136.4 C 223kPa 7:10
VENTING 7:11
AIR DRYING 7:59
CYCLE COMPLETE 8:04

Digital Signature #
8922E8ECC774636E
