
Midmark M11 - v1.0.0
Total Cycles: 695

Name: MIDMARK
Sterilizer ID: V2178198
02/24/2021 07:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.8 F 0.0
2:00 77.8 F 0.0
4:00 79.7 F 0.0
6:00 102.1 F 0.1
8:00 145.7 F 0.2
10:00 186.0 F 1.6
12:00 203.0 F 5.0
14:00 222.2 F 8.2
16:00 230.1 F 12.0
18:00 233.6 F 7.6
20:00 245.3 F 13.0
22:00 241.2 F 10.7
24:00 237.2 F 8.9
26:00 252.3 F 16.2
28:00 264.1 F 23.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.3 F 28.4
2:00 272.1 F 28.7
3:00 271.8 F 28.0
4:00 272.7 F 28.8

Min 271.2 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.7 F 0.0
5:00 239.6 F 0.0
10:00 226.5 F 0.0
15:00 230.1 F 0.0
20:00 231.3 F 0.0
25:00 210.8 F 0.0
30:00 201.3 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 29:57
STERILIZING: 4:00
VENTING: 2:54
DRYING: 30:00
TOTAL CYCLE: 01:07:46
CYCLE COMPLETE

02/24/2021 08:30 AM
ACCEPTED: X
REJECTED:
Operator: U01