
Midmark M11 - v1.0.0
Total Cycles: 3738

Name: RITTER 1
Sterilizer ID: V2178198
09/08/2023 07:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.4 F 0.0
2:00 73.8 F 0.0
4:00 87.5 F 0.0
6:00 117.6 F 0.3
8:00 143.0 F 1.7
10:00 171.5 F 3.4
12:00 185.7 F 5.8
14:00 198.2 F 9.0
16:00 212.0 F 12.6
18:00 216.4 F 5.0
20:00 227.7 F 8.8
22:00 238.0 F 13.0
24:00 236.2 F 9.6
26:00 239.1 F 9.8
28:00 246.7 F 13.5
30:00 257.9 F 19.5
32:00 267.9 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.6 F 28.7
2:00 271.6 F 28.3
3:00 272.5 F 28.7
4:00 271.4 F 28.2

Min 271.4 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 218.5 F 0.0
10:00 209.8 F 0.0
15:00 215.7 F 0.0
20:00 196.5 F 0.0
25:00 189.8 F 0.0
30:00 186.3 F 0.0

Drying Complete

FILLING: 1:04
HEATING: 33:22
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:11:52
CYCLE COMPLETE

09/08/2023 09:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____