
Midmark M11 - v1.0.5
Total Cycles: 1740

Name: RITTER2
Sterilizer ID: V2313228
11/17/2022 09:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.4 F 0.0
2:00 174.3 F 0.1
4:00 184.9 F 0.7
6:00 215.0 F 4.6
8:00 228.9 F 9.7
10:00 236.1 F 8.6
12:00 239.5 F 12.0
14:00 238.7 F 10.4
16:00 237.8 F 9.5
18:00 252.3 F 16.7
20:00 264.3 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.7
1:00 272.4 F 29.4
2:00 271.5 F 28.7
3:00 271.8 F 29.4
4:00 272.2 F 29.2

Min 271.0 F 28.5
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.7 F 0.0
5:00 208.9 F 0.0
10:00 220.2 F 0.0
15:00 230.7 F 0.0
20:00 228.3 F 0.0
25:00 227.7 F 0.0
30:00 226.9 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 21:58
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:10
CYCLE COMPLETE

11/17/2022 10:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____