

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.6uS / 0.3ppm  
CYCLE NUMBER 000930  
16:41 11/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:44  
PRESSURIZING 2:58  
STERILIZING 4:22  
136.0 C 228kPa 4:22  
136.2 C 230kPa 4:52  
136.4 C 233kPa 5:22  
136.3 C 233kPa 5:52  
136.0 C 228kPa 6:22  
136.3 C 232kPa 6:52  
136.4 C 233kPa 7:22  
136.3 C 232kPa 7:52  
136.3 C 232kPa 7:52  
VENTING 7:52  
AIR DRYING 8:47  
CYCLE COMPLETE 23:47  
-----

Digital Signature #  
99E676BDA65D8B5E  
-----