

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3351

Name: RITTER 1  
Sterilizer ID: V2178198  
06/07/2023 12:39 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 159.9 F 0.0  
2:00 171.1 F 0.0  
4:00 181.7 F 1.5  
6:00 200.8 F 6.0  
8:00 216.1 F 10.7  
10:00 223.8 F 6.2  
12:00 238.6 F 11.5  
14:00 235.3 F 9.0  
16:00 239.6 F 9.7  
18:00 248.2 F 14.4  
20:00 260.3 F 21.3  
22:00 271.5 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.5 F 29.2  
2:00 271.5 F 28.5  
3:00 272.5 F 29.4  
4:00 272.1 F 28.9

Min 271.5 F 28.5  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.6 F 0.0  
5:00 213.5 F 0.0  
10:00 222.1 F 0.0  
15:00 234.6 F 0.0  
20:00 230.5 F 0.0  
25:00 224.4 F 0.0  
30:00 223.6 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 22:38  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:00:54  
CYCLE COMPLETE

06/07/2023 01:39 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_