
Midmark M11 - v1.0.0
Total Cycles: 1482

Name: RITTER 1
Sterilizer ID: V2178198
12/21/2021 12:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.5 F 0.0
2:00 195.5 F 3.3
4:00 195.6 F 7.7
6:00 222.3 F 13.5
8:00 228.7 F 8.3
10:00 239.7 F 12.1
12:00 239.8 F 11.2
14:00 237.9 F 9.4
16:00 251.4 F 16.1
18:00 262.9 F 22.7
20:00 272.5 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.4 F 29.3
2:00 272.6 F 29.0
3:00 271.6 F 28.4
4:00 271.5 F 28.7

Min 271.2 F 28.3
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.2 F 0.0
5:00 232.1 F 0.0
10:00 233.4 F 0.0
15:00 229.7 F 0.0
20:00 229.9 F 0.0
25:00 229.3 F 0.0
30:00 230.8 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 20:18
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:20
CYCLE COMPLETE

12/21/2021 01:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____