
Midmark M11 - v1.0.5
Total Cycles: 3367

Name: RITTER2
Sterilizer ID: V2313228
02/07/2024 05:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 113.0 F 0.0
2:00 122.4 F 0.6
4:00 145.2 F 3.1
6:00 174.1 F 7.1
8:00 196.0 F 11.7
10:00 217.0 F 4.9
12:00 232.3 F 9.3
14:00 234.7 F 10.1
16:00 241.6 F 11.9
18:00 240.6 F 10.8
20:00 254.0 F 17.6
22:00 265.9 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.4 F 29.1
2:00 272.0 F 29.2
3:00 272.3 F 29.0
4:00 271.8 F 29.1

Min 271.4 F 28.6
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 170.6 F 0.0
5:00 235.5 F 0.0
10:00 205.3 F 0.0
15:00 199.7 F 0.0
20:00 196.6 F 0.0
25:00 194.4 F 0.0
30:00 192.3 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 23:37
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:02:08
CYCLE COMPLETE

02/07/2024 06:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____