
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 004697
10:06 14/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:41
STERILIZING 3:32
136.0 C 220kPa 3:32
136.4 C 225kPa 4:02
136.2 C 219kPa 4:32
136.4 C 221kPa 5:02
136.4 C 223kPa 5:32
136.4 C 223kPa 6:02
136.2 C 221kPa 6:32
136.4 C 223kPa 7:02
136.4 C 223kPa 7:02
VENTING 7:02
AIR DRYING 7:49
CYCLE COMPLETE 7:57

Digital Signature #
CA34AC3677ACEFD6
