
Midmark M11 - v1.0.5
Total Cycles: 4219

Name: RITTER2
Sterilizer ID: V2313228
12/16/2024 04:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 117.2 F 0.0
2:00 119.0 F 0.6
4:00 125.4 F 2.7
6:00 150.9 F 5.7
8:00 174.1 F 9.3
10:00 195.8 F 13.1
12:00 210.0 F 6.3
14:00 224.2 F 10.5
16:00 231.2 F 7.8
18:00 239.1 F 11.7
20:00 237.7 F 9.9
22:00 250.2 F 16.0
24:00 261.1 F 22.2
26:00 271.3 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.2
1:00 272.1 F 29.8
2:00 272.4 F 29.5
3:00 271.5 F 29.4
4:00 272.7 F 29.9

Min 271.4 F 29.1
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.2 F 0.0
5:00 206.4 F 0.0
10:00 212.1 F 0.0
15:00 213.8 F 0.0
20:00 217.0 F 0.0
25:00 231.5 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 26:34
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:05:23
CYCLE COMPLETE

12/16/2024 05:05 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____