

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1821

Name: RITTER 1  
Sterilizer ID: V2178198  
04/13/2022 01:46 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.6 F 0.0  
2:00 159.5 F 2.9  
4:00 183.9 F 7.2  
6:00 204.7 F 12.1  
8:00 222.3 F 6.0  
10:00 231.2 F 11.1  
12:00 231.4 F 7.5  
14:00 246.1 F 13.5  
16:00 243.4 F 12.0  
18:00 256.1 F 18.6  
20:00 267.7 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.7  
1:00 271.7 F 28.5  
2:00 272.2 F 28.7  
3:00 272.5 F 29.0  
4:00 272.3 F 29.2

Min 271.4 F 28.5  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.8 F 0.0  
5:00 228.0 F 0.0  
10:00 227.6 F 0.0  
15:00 228.0 F 0.0  
20:00 229.9 F 0.0  
25:00 228.5 F 0.0  
30:00 229.2 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 21:55  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:00:08  
CYCLE COMPLETE

04/13/2022 02:46 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_