
Midmark M11 - v1.0.0
Total Cycles: 4538

Name: RITTER 1
Sterilizer ID: V2178198
05/09/2024 09:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.0 F 0.0
2:00 183.6 F 2.8
4:00 183.6 F 6.7
6:00 201.8 F 11.3
8:00 215.8 F 5.5
10:00 230.0 F 10.5
12:00 231.6 F 7.3
14:00 242.3 F 12.8
16:00 240.8 F 11.2
18:00 252.7 F 17.1
20:00 263.1 F 23.2
22:00 272.5 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.9 F 29.6
2:00 272.0 F 29.0
3:00 271.4 F 28.9
4:00 272.5 F 29.6

Min 271.2 F 28.7
Max 273.0 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 209.5 F 0.0
10:00 200.3 F 0.0
15:00 224.5 F 0.0
20:00 236.7 F 0.0
25:00 224.1 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 22:27
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:01:27
CYCLE COMPLETE

05/09/2024 10:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____