

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5435

Name: RITTER 1  
Sterilizer ID: V2178198  
03/24/2025 11:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.3 F 0.0  
2:00 160.1 F 2.7  
4:00 171.4 F 6.1  
6:00 193.5 F 10.5  
8:00 219.2 F 3.9  
10:00 228.1 F 8.2  
12:00 238.7 F 13.1  
14:00 238.1 F 10.2  
16:00 235.8 F 8.2  
18:00 248.9 F 14.9  
20:00 261.0 F 21.6  
22:00 271.6 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.8 F 29.6  
2:00 271.7 F 28.8  
3:00 272.3 F 29.6  
4:00 272.6 F 29.4

Min 271.2 F 28.7  
Max 273.0 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 175.6 F 0.0  
5:00 225.9 F 0.0  
10:00 231.3 F 0.0  
15:00 224.8 F 0.0  
20:00 219.7 F 0.0  
25:00 224.8 F 0.0  
30:00 233.5 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 23:28  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:02:16  
CYCLE COMPLETE

03/24/2025 12:27 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_