

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1972

Name: RITTER 1  
Sterilizer ID: V2178198  
05/30/2022 11:18 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.3 F 0.0  
2:00 194.0 F 4.5  
4:00 198.7 F 9.4  
6:00 224.1 F 7.5  
8:00 229.6 F 9.7  
10:00 233.5 F 8.1  
12:00 242.7 F 12.3  
14:00 240.8 F 10.6  
16:00 254.3 F 17.4  
18:00 266.0 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.2  
1:00 272.2 F 28.9  
2:00 271.3 F 28.1  
3:00 272.1 F 29.0  
4:00 271.8 F 28.5  
  
Min 271.1 F 28.1  
Max 272.4 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.1 F 0.0  
5:00 215.4 F 0.0  
10:00 234.3 F 0.0  
15:00 239.4 F 0.0  
20:00 227.9 F 0.0  
25:00 225.6 F 0.0  
30:00 223.2 F 0.0  
Drying Complete

FILLING: 1:31  
HEATING: 19:40  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:20  
CYCLE COMPLETE

05/30/2022 12:16 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_