
Midmark M11 - v1.0.5
Total Cycles: 1659

Name: RITTER2
Sterilizer ID: V2313228
10/26/2022 10:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.7 F 0.0
2:00 171.8 F 2.8
4:00 193.3 F 7.2
6:00 208.5 F 12.5
8:00 221.2 F 6.9
10:00 231.3 F 12.3
12:00 234.0 F 9.5
14:00 238.7 F 9.4
16:00 246.9 F 14.6
18:00 259.1 F 21.3
20:00 269.9 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.5
1:00 272.3 F 30.4
2:00 272.5 F 30.0
3:00 271.6 F 29.5
4:00 271.2 F 29.7

Min 271.0 F 29.3
Max 272.9 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 208.9 F 0.0
10:00 230.7 F 0.0
15:00 231.0 F 0.0
20:00 227.6 F 0.0
25:00 226.0 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 21:03
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:17
CYCLE COMPLETE

10/26/2022 11:21 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____