

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4478

Name: RITTER2  
Sterilizer ID: V2313228  
04/02/2025 12:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.6 F 0.0  
2:00 161.3 F 2.4  
4:00 166.5 F 6.2  
6:00 188.1 F 10.8  
8:00 222.5 F 4.7  
10:00 223.7 F 9.6  
12:00 235.8 F 9.8  
14:00 239.7 F 11.8  
16:00 238.0 F 9.9  
18:00 250.0 F 15.8  
20:00 260.5 F 21.9  
22:00 270.3 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.5  
1:00 271.7 F 30.3  
2:00 272.4 F 30.6  
3:00 272.3 F 30.4  
4:00 272.2 F 30.2

Min 270.9 F 29.3  
Max 273.1 F 30.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.4 F 0.0  
5:00 192.8 F 0.0  
10:00 205.2 F 0.0  
15:00 212.3 F 0.0  
20:00 207.4 F 0.0  
25:00 211.9 F 0.0  
30:00 201.3 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 23:01  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:01:58  
CYCLE COMPLETE

04/02/2025 01:16 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_