
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000482
11:06 19/03/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:44
PRESSURIZING 3:48
STERILIZING 4:54
136.0 C 226kPa 4:54
136.4 C 231kPa 5:24
136.1 C 230kPa 5:54
136.1 C 230kPa 6:24
136.3 C 232kPa 6:54
136.4 C 233kPa 7:24
136.4 C 232kPa 7:54
136.1 C 231kPa 8:24
136.1 C 231kPa 8:24
VENTING 8:24
AIR DRYING 9:19
CYCLE COMPLETE 24:19

Digital Signature #
5F8BF4641B3FEA5F
