
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001028
9:59 08/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:49
PRESSURIZING 2:56
STERILIZING 4:01
136.3 C 221kPa 4:01
136.4 C 225kPa 4:31
136.4 C 223kPa 5:01
136.1 C 222kPa 5:31
136.2 C 222kPa 6:01
136.0 C 223kPa 6:31
136.4 C 223kPa 7:01
136.2 C 221kPa 7:31
136.2 C 221kPa 7:31
VENTING 7:31
AIR DRYING 8:21
CYCLE COMPLETE 23:21

Digital Signature #
5DC98FACF95E6936
