

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
1.1uS / 0.7ppm  
CYCLE NUMBER 001492  
14:13 10/12/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 5:00  
PRESSURIZING 6:55  
STERILIZING 8:28  
136.1 C 220kPa 8:28  
136.4 C 222kPa 8:59  
136.4 C 222kPa 9:29  
136.4 C 222kPa 9:59  
136.5 C 220kPa 10:29  
136.4 C 222kPa 10:59  
136.2 C 221kPa 11:29  
136.0 C 221kPa 11:59  
VENTING 11:59  
AIR DRYING 13:05  
CYCLE COMPLETE 27:30  
-----

Digital Signature #  
20F6115148D05293  
-----