

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001304  
15:08 19/10/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:58  
PRESSURIZING 2:07  
STERILIZING 3:14  
136.0 C 220kPa 3:14  
136.1 C 220kPa 3:44  
136.2 C 223kPa 4:14  
136.1 C 221kPa 4:44  
136.4 C 224kPa 5:14  
136.4 C 224kPa 5:44  
136.3 C 224kPa 6:14  
136.2 C 222kPa 6:44  
136.2 C 222kPa 6:44  
VENTING 6:44  
AIR DRYING 7:35  
CYCLE COMPLETE 22:35  
-----

Digital Signature #  
4C1D0DFC1D8C0C58  
-----