

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2634

Name: RITTER2  
Sterilizer ID: V2313228  
07/12/2023 08:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.8 F 0.0  
2:00 73.5 F 0.0  
4:00 89.1 F 0.0  
6:00 102.6 F 0.0  
8:00 147.5 F 0.1  
10:00 173.5 F 3.2  
12:00 187.9 F 6.7  
14:00 202.4 F 10.4  
16:00 218.0 F 7.9  
18:00 226.5 F 8.0  
20:00 236.0 F 12.2  
22:00 234.8 F 9.1  
24:00 239.9 F 10.6  
26:00 245.4 F 13.3  
28:00 256.6 F 19.4  
30:00 267.0 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.0  
1:00 272.7 F 29.8  
2:00 271.9 F 29.0  
3:00 271.7 F 29.3  
4:00 272.9 F 29.7

Min 271.3 F 28.8  
Max 273.0 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.3 F 0.0  
5:00 201.5 F 0.0  
10:00 217.5 F 0.0  
15:00 220.9 F 0.0  
20:00 215.7 F 0.0  
25:00 225.0 F 0.0  
30:00 226.8 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 31:36  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:09:34  
CYCLE COMPLETE

07/12/2023 09:21 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_