
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 000076
14:28 07/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:16
STERILIZING 4:31
136.4 C 225kPa 4:31
136.4 C 225kPa 5:01
136.4 C 221kPa 5:31
136.4 C 225kPa 6:01
136.4 C 225kPa 6:31
136.3 C 224kPa 7:01
136.4 C 226kPa 7:31
136.3 C 224kPa 8:01
136.3 C 224kPa 8:01
VENTING 8:01
AIR DRYING 8:52
CYCLE COMPLETE 9:14

Digital Signature #
71219406EC763E16
