
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002012
10:50 11/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:39
PRESSURIZING 2:47
STERILIZING 3:50
136.4 C 224kPa 3:50
136.1 C 222kPa 4:20
136.4 C 225kPa 4:50
136.4 C 225kPa 5:20
136.2 C 222kPa 5:50
136.2 C 223kPa 6:20
136.4 C 222kPa 6:50
136.3 C 224kPa 7:20
136.3 C 224kPa 7:20
VENTING 7:20
AIR DRYING 8:13
CYCLE COMPLETE 38:13

Digital Signature #
F563C7562B97B0C6
