
Midmark M11 - v1.0.5
Total Cycles: 2540

Name: RITTER2
Sterilizer ID: V2313228
06/20/2023 02:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.2 F 0.0
2:00 170.2 F 3.5
4:00 180.7 F 7.5
6:00 202.1 F 12.5
8:00 219.8 F 7.0
10:00 232.7 F 12.1
12:00 234.1 F 9.3
14:00 238.2 F 9.7
16:00 246.4 F 14.0
18:00 258.3 F 20.6
20:00 269.3 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 273.1 F 29.9
2:00 272.0 F 29.2
3:00 271.7 F 29.6
4:00 272.9 F 29.9

Min 271.4 F 29.0
Max 273.2 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 223.7 F 0.0
10:00 224.9 F 0.0
15:00 227.1 F 0.0
20:00 222.9 F 0.0
25:00 221.6 F 0.0
30:00 221.1 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 21:07
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:43
CYCLE COMPLETE

06/20/2023 03:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____