
Midmark M11 - v1.0.0
Total Cycles: 1659

Name: RITTER 1
Sterilizer ID: V2178198
02/23/2022 10:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.0 F 0.0
2:00 184.3 F 2.6
4:00 187.9 F 6.6
6:00 209.7 F 12.4
8:00 223.7 F 6.7
10:00 237.0 F 12.1
12:00 236.8 F 9.5
14:00 238.9 F 9.0
16:00 249.1 F 14.7
18:00 261.1 F 21.4
20:00 272.2 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.6 F 28.8
2:00 271.5 F 28.1
3:00 272.8 F 28.9
4:00 271.8 F 28.2

Min 271.5 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.5 F 0.0
5:00 224.6 F 0.0
10:00 237.3 F 0.0
15:00 226.8 F 0.0
20:00 227.2 F 0.0
25:00 227.4 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 20:30
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:11
CYCLE COMPLETE

02/23/2022 11:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____