
Midmark M11 - v1.0.0
Total Cycles: 4126

Name: RITTER 1
Sterilizer ID: V2178198
12/15/2023 02:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.5 F 0.0
2:00 167.4 F 2.7
4:00 187.0 F 6.3
6:00 208.6 F 11.2
8:00 225.3 F 5.1
10:00 231.5 F 9.9
12:00 236.0 F 9.0
14:00 242.3 F 12.0
16:00 240.5 F 10.4
18:00 252.7 F 16.6
20:00 264.0 F 23.0
22:00 271.2 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 271.8 F 28.4
2:00 272.3 F 28.6
3:00 272.3 F 28.9
4:00 272.1 F 29.0

Min 271.2 F 28.4
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.3 F 0.0
5:00 217.1 F 0.0
10:00 213.3 F 0.0
15:00 195.9 F 0.0
20:00 191.0 F 0.0
25:00 187.9 F 0.0
30:00 187.2 F 0.0
Drying Complete

FILLING: 1:52
HEATING: 22:47
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:02:21
CYCLE COMPLETE

12/15/2023 03:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____