
Midmark M11 - v1.0.0
Total Cycles: 3781

Name: RITTER 1
Sterilizer ID: V2178198
09/15/2023 12:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.1 F 0.0
2:00 146.7 F 0.0
4:00 169.9 F 2.9
6:00 194.3 F 6.7
8:00 214.6 F 11.3
10:00 224.1 F 5.9
12:00 234.9 F 10.9
14:00 233.3 F 7.4
16:00 246.1 F 12.9
18:00 244.2 F 12.2
20:00 256.8 F 18.7
22:00 267.5 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.2 F 28.4
2:00 272.4 F 28.8
3:00 271.5 F 28.1
4:00 272.7 F 28.8

Min 271.3 F 28.0
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 163.8 F 0.0
5:00 194.0 F 0.0
10:00 236.1 F 0.0
15:00 210.0 F 0.0
20:00 232.0 F 0.0
25:00 190.1 F 0.0
30:00 188.9 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 23:21
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:01:35
CYCLE COMPLETE

09/15/2023 01:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____