
Midmark M11 - v1.0.5
Total Cycles: 2676

Name: RITTER2
Sterilizer ID: V2313228
07/20/2023 01:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.2 F 0.0
2:00 147.7 F 0.4
4:00 163.0 F 3.7
6:00 191.7 F 8.2
8:00 211.9 F 12.7
10:00 220.1 F 7.5
12:00 235.4 F 12.6
14:00 235.0 F 9.8
16:00 236.2 F 9.0
18:00 246.9 F 14.4
20:00 258.8 F 20.9
22:00 269.1 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.1 F 30.2
2:00 272.1 F 29.7
3:00 272.5 F 30.1
4:00 272.0 F 30.2

Min 270.8 F 29.1
Max 273.1 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.0 F 0.0
5:00 217.8 F 0.0
10:00 222.6 F 0.0
15:00 222.4 F 0.0
20:00 225.3 F 0.0
25:00 226.8 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 22:58
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:01:14
CYCLE COMPLETE

07/20/2023 02:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____