
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001467
13:09 20/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
CYCLE ABORTED 2:36
WATER LEVEL LOW
NOT STERILE
END TIME 2:36

Digital Signature #
EC1E32A38C096172
