

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2205

Name: RITTER2  
Sterilizer ID: V2313228  
04/03/2023 09:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 77.2 F 0.0  
2:00 77.9 F 0.0  
4:00 83.2 F 0.0  
6:00 103.8 F 1.2  
8:00 161.1 F 4.2  
10:00 186.3 F 7.9  
12:00 195.0 F 12.0  
14:00 215.0 F 4.5  
16:00 230.1 F 8.3  
18:00 239.0 F 12.3  
20:00 232.0 F 8.7  
22:00 241.3 F 11.8  
24:00 244.2 F 12.6  
26:00 255.7 F 18.7  
28:00 266.1 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.8  
1:00 271.9 F 29.0  
2:00 272.4 F 29.5  
3:00 271.1 F 29.1  
4:00 271.6 F 28.9

Min 271.0 F 28.6  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.1 F 0.0  
5:00 237.8 F 0.0  
10:00 223.5 F 0.0  
15:00 202.6 F 0.0  
20:00 189.8 F 0.0  
25:00 190.7 F 0.0  
30:00 191.1 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 29:40  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:07:37  
CYCLE COMPLETE

04/03/2023 10:28 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_