
Midmark M11 - v1.0.5
Total Cycles: 606

Name: RITTER2
Sterilizer ID: V2313228
12/06/2021 03:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.8 F 0.0
2:00 161.2 F 2.9
4:00 189.3 F 7.8
6:00 207.8 F 13.5
8:00 224.3 F 8.2
10:00 237.5 F 12.0
12:00 240.3 F 12.1
14:00 240.9 F 11.5
16:00 254.5 F 18.5
18:00 266.1 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.3
1:00 272.2 F 29.5
2:00 272.7 F 30.0
3:00 271.8 F 29.7
4:00 272.2 F 29.5

Min 271.6 F 29.2
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.2 F 0.0
5:00 219.4 F 0.0
10:00 221.4 F 0.0
15:00 221.5 F 0.0
20:00 222.6 F 0.0
25:00 224.9 F 0.0
30:00 224.7 F 0.0

Drying Complete

FILLING: 2:18
HEATING: 19:35
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:58
CYCLE COMPLETE

12/06/2021 04:47 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____