

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000204  
13:51 21/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:27  
PRESSURIZING 3:33  
STERILIZING 5:04  
136.0 C 229kPa 5:04  
136.3 C 233kPa 5:34  
136.4 C 233kPa 6:04  
136.3 C 232kPa 6:34  
136.4 C 233kPa 7:04  
136.3 C 232kPa 7:34  
136.4 C 233kPa 8:04  
136.4 C 233kPa 8:34  
136.4 C 233kPa 8:34  
VENTING 8:34  
AIR DRYING 9:45  
CYCLE COMPLETE 9:54  
-----

Digital Signature #  
1A12EEB4CAFDD0DE  
-----