

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
1.1uS / 0.7ppm  
CYCLE NUMBER 001104  
9:53 03/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:46  
PRESSURIZING 2:54  
STERILIZING 4:03  
136.0 C 220kPa 4:03  
136.4 C 224kPa 4:33  
136.4 C 224kPa 5:03  
136.4 C 225kPa 5:33  
136.4 C 222kPa 6:03  
136.4 C 221kPa 6:33  
136.3 C 223kPa 7:03  
136.4 C 225kPa 7:33  
136.4 C 225kPa 7:33  
VENTING 7:33  
AIR DRYING 8:22  
CYCLE COMPLETE 14:38

-----  
Digital Signature #  
3C8C339A133CA6D5  
-----