
Midmark M11 - v1.0.5
Total Cycles: 2597

Name: RITTER2
Sterilizer ID: V2313228
06/30/2023 02:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.7 F 0.0
2:00 194.8 F 3.9
4:00 205.9 F 9.6
6:00 221.9 F 7.2
8:00 235.9 F 11.7
10:00 235.8 F 9.7
12:00 238.1 F 9.0
14:00 249.9 F 15.6
16:00 262.0 F 22.6
18:00 272.8 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.5 F 29.6
2:00 271.3 F 28.9
3:00 272.4 F 29.8
4:00 272.1 F 29.3

Min 271.1 F 28.9
Max 272.5 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.4 F 0.0
5:00 220.5 F 0.0
10:00 230.7 F 0.0
15:00 229.7 F 0.0
20:00 228.6 F 0.0
25:00 228.5 F 0.0
30:00 227.2 F 0.0

Drying Complete

FILLING: 2:20
HEATING: 18:20
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:57:57
CYCLE COMPLETE

06/30/2023 03:02 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____