
Midmark M11 - v1.0.5
Total Cycles: 3870

Name: RITTER2
Sterilizer ID: V2313228
08/16/2024 04:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.7 F 0.0
2:00 150.7 F 2.6
4:00 178.2 F 7.1
6:00 205.2 F 12.9
8:00 220.8 F 6.9
10:00 234.8 F 12.2
12:00 236.1 F 9.7
14:00 237.0 F 8.9
16:00 249.8 F 15.6
18:00 262.5 F 22.8
20:00 272.2 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.8
1:00 272.6 F 29.3
2:00 272.3 F 29.4
3:00 272.2 F 28.9
4:00 272.7 F 29.7

Min 271.7 F 28.7
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 169.9 F 0.0
5:00 196.0 F 0.0
10:00 206.8 F 0.0
15:00 213.3 F 0.0
20:00 223.2 F 0.0
25:00 224.7 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 1:59
HEATING: 20:10
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:18
CYCLE COMPLETE

08/16/2024 05:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____