
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000008
15:07 13/12/2019
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 3:03
CYCLE FAULT 15
NOT STERILE
END TIME 3:03

Digital Signature #
C81EE58C0EA7ECF0
