
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001697
16:47 05/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:26
PRESSURIZING 1:31
STERILIZING 2:30
136.0 C 220kPa 2:30
136.4 C 224kPa 3:00
136.2 C 222kPa 3:30
136.4 C 225kPa 4:00
136.4 C 225kPa 4:30
136.4 C 225kPa 5:00
136.4 C 225kPa 5:30
136.1 C 223kPa 6:00
136.1 C 223kPa 6:00
VENTING 6:00
AIR DRYING 6:50
CYCLE COMPLETE 6:56

Digital Signature #
3EAF23E367ABBBFE
