

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2531

Name: RITTER2  
Sterilizer ID: V2313228  
06/19/2023 12:58 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.1 F 0.0  
2:00 193.5 F 4.0  
4:00 198.2 F 8.9  
6:00 217.7 F 7.9  
8:00 231.4 F 10.3  
10:00 230.8 F 7.9  
12:00 241.4 F 12.0  
14:00 244.4 F 13.1  
16:00 257.3 F 20.0  
18:00 268.6 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.2 F 29.6  
2:00 271.8 F 29.6  
3:00 272.1 F 29.4  
4:00 271.7 F 29.6  
  
Min 271.4 F 29.0  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.4 F 0.0  
5:00 223.6 F 0.0  
10:00 233.8 F 0.0  
15:00 225.7 F 0.0  
20:00 231.0 F 0.0  
25:00 229.7 F 0.0  
30:00 230.4 F 0.0  
  
Drying Complete

FILLING: 1:60  
HEATING: 19:07  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:22  
CYCLE COMPLETE

06/19/2023 01:56 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_