
Midmark M11 - v1.0.0
Total Cycles: 1039

Name: RITTER 1
Sterilizer ID: V2178198
07/21/2021 12:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.2 F 0.0
2:00 167.7 F 2.6
4:00 186.1 F 6.0
6:00 207.3 F 10.8
8:00 218.3 F 5.0
10:00 233.5 F 10.0
12:00 233.7 F 8.3
14:00 243.0 F 12.5
16:00 241.4 F 10.9
18:00 254.3 F 17.4
20:00 265.5 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.8 F 28.9
2:00 271.8 F 28.2
3:00 272.2 F 28.9
4:00 272.7 F 28.8

Min 271.5 F 28.1
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.3 F 0.0
5:00 220.9 F 0.0
10:00 204.7 F 0.0
15:00 201.0 F 0.0
20:00 196.3 F 0.0
25:00 192.5 F 0.0
30:00 188.7 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 21:47
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:18
CYCLE COMPLETE

07/21/2021 01:54 PM
ACCEPTED: X
REJECTED:
Operator: U01