
Midmark M11 - v1.0.5
Total Cycles: 2519

Name: RITTER2
Sterilizer ID: V2313228
06/15/2023 08:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.5 F 0.0
2:00 160.6 F 2.2
4:00 169.9 F 5.3
6:00 189.0 F 9.6
8:00 211.0 F 6.7
10:00 225.1 F 8.4
12:00 236.3 F 13.4
14:00 235.6 F 10.3
16:00 237.0 F 8.4
18:00 247.9 F 14.9
20:00 259.4 F 21.4
22:00 270.4 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.3
1:00 273.0 F 29.9
2:00 271.7 F 29.2
3:00 273.1 F 30.1
4:00 272.4 F 29.5

Min 271.6 F 29.2
Max 273.2 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.3 F 0.0
5:00 228.2 F 0.0
10:00 222.2 F 0.0
15:00 226.7 F 0.0
20:00 227.0 F 0.0
25:00 224.4 F 0.0
30:00 219.9 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 22:55
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:01:02
CYCLE COMPLETE

06/15/2023 09:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____