
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 005049
16:56 01/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:01
STERILIZING 2:59
136.5 C 220kPa 2:59
136.3 C 223kPa 3:29
136.4 C 223kPa 3:59
136.1 C 219kPa 4:29
136.4 C 223kPa 4:59
136.4 C 223kPa 5:29
136.4 C 225kPa 5:59
135.9 C 217kPa 6:29
135.9 C 217kPa 6:29
VENTING 6:30
AIR DRYING 7:21
CYCLE COMPLETE 20:06

Digital Signature #
B6814AC01BC9EAD7
