

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2145

Name: RITTER2  
Sterilizer ID: V2313228  
03/14/2023 09:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 151.7 F 0.0  
2:00 148.8 F 1.7  
4:00 154.0 F 4.5  
6:00 181.8 F 8.8  
8:00 205.9 F 9.1  
10:00 222.5 F 8.3  
12:00 233.6 F 12.6  
14:00 237.3 F 10.9  
16:00 236.7 F 9.5  
18:00 250.6 F 16.3  
20:00 262.7 F 23.3  
22:00 272.4 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.6  
1:00 272.7 F 29.9  
2:00 272.2 F 30.0  
3:00 272.5 F 29.9  
4:00 272.0 F 29.9

Min 271.7 F 29.4  
Max 273.2 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.4 F 0.0  
5:00 212.0 F 0.0  
10:00 213.5 F 0.0  
15:00 212.9 F 0.0  
20:00 217.2 F 0.0  
25:00 233.7 F 0.0  
30:00 218.1 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 22:59  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:01:15  
CYCLE COMPLETE

03/14/2023 10:42 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_