
Midmark M11 - v1.0.0
Total Cycles: 2875

Name: RITTER 1
Sterilizer ID: V2178198
02/08/2023 12:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.7 F 0.0
2:00 191.6 F 4.6
4:00 198.6 F 9.5
6:00 217.7 F 7.3
8:00 229.9 F 9.4
10:00 237.2 F 9.1
12:00 242.6 F 12.2
14:00 241.0 F 10.7
16:00 254.1 F 17.5
18:00 265.5 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.4 F 29.1
2:00 271.8 F 28.4
3:00 271.8 F 28.9
4:00 272.2 F 29.0

Min 271.2 F 28.2
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.7 F 0.0
5:00 225.9 F 0.0
10:00 236.4 F 0.0
15:00 231.6 F 0.0
20:00 225.8 F 0.0
25:00 226.5 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 19:42
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:29
CYCLE COMPLETE

02/08/2023 01:06 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____