
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007221
6:43 18/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 0:50
PRESSURIZING 0:57
STERILIZING 2:07
136.5 C 221kPa 2:07
136.3 C 223kPa 2:37
136.4 C 220kPa 3:07
136.4 C 222kPa 3:37
136.4 C 218kPa 4:07
136.4 C 221kPa 4:37
136.4 C 219kPa 5:07
136.1 C 221kPa 5:37
136.1 C 221kPa 5:37
VENTING 5:37
AIR DRYING 6:26
CYCLE COMPLETE 6:46

Digital Signature #
DBA5CCF00078A24F
