
Midmark M11 - v1.0.0
Total Cycles: 3768

Name: RITTER 1
Sterilizer ID: V2178198
09/13/2023 03:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.9 F 0.0
2:00 178.9 F 2.8
4:00 202.2 F 8.1
6:00 218.4 F 9.6
8:00 234.0 F 9.5
10:00 235.3 F 8.8
12:00 245.5 F 13.2
14:00 244.1 F 12.1
16:00 257.8 F 19.3
18:00 269.7 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.1
1:00 272.5 F 28.5
2:00 272.6 F 28.9
3:00 271.5 F 28.0
4:00 272.4 F 28.5

Min 271.4 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.1 F 0.0
5:00 210.8 F 0.0
10:00 244.6 F 0.0
15:00 225.0 F 0.0
20:00 226.3 F 0.0
25:00 229.5 F 0.0
30:00 223.3 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 18:45
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:41
CYCLE COMPLETE

09/13/2023 04:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____