
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001128
15:23 20/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:18
PRESSURIZING 4:18
STERILIZING 5:23
136.4 C 221kPa 5:23
136.3 C 225kPa 5:53
136.3 C 225kPa 6:23
136.3 C 224kPa 6:53
136.4 C 225kPa 7:23
136.4 C 226kPa 7:53
136.3 C 226kPa 8:23
136.4 C 225kPa 8:53
136.4 C 225kPa 8:53
VENTING 8:53
AIR DRYING 9:51
CYCLE COMPLETE 21:42

Digital Signature #
0854B00DB1030D04
