

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000001  
16:47 29/11/2019  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:41  
PRESSURIZING 2:46  
STERILIZING 3:34  
136.4 C 228kPa 3:34  
136.4 C 227kPa 4:04  
135.9 C 221kPa 4:34  
136.0 C 224kPa 5:04  
136.3 C 224kPa 5:34  
136.4 C 226kPa 6:04  
136.4 C 223kPa 6:34  
136.4 C 224kPa 7:04  
136.4 C 224kPa 7:04  
VENTING 7:04  
AIR DRYING 7:52  
CYCLE COMPLETE 37:52

-----  
Digital Signature #  
0B986FA7090FDDB0  
-----