
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001761
15:09 01/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:35
PRESSURIZING 2:39
STERILIZING 3:47
136.5 C 226kPa 3:47
136.2 C 222kPa 4:17
136.4 C 224kPa 4:47
136.4 C 224kPa 5:17
136.4 C 224kPa 5:47
136.4 C 226kPa 6:17
136.4 C 223kPa 6:47
136.4 C 221kPa 7:17
136.4 C 221kPa 7:17
VENTING 7:17
AIR DRYING 8:10
CYCLE COMPLETE 20:46

Digital Signature #
7099B8EED1F6ACBD
