

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4045

Name: RITTER 1  
Sterilizer ID: V2178198  
11/21/2023 09:24 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.3 F 0.0  
2:00 78.4 F 0.0  
4:00 84.7 F 0.6  
6:00 103.5 F 2.7  
8:00 134.3 F 5.2  
10:00 162.6 F 8.3  
12:00 183.2 F 11.7  
14:00 204.4 F 2.7  
16:00 212.5 F 5.9  
18:00 222.4 F 9.3  
20:00 232.2 F 13.1  
22:00 233.6 F 8.5  
24:00 244.4 F 13.4  
26:00 242.3 F 11.3  
28:00 253.9 F 17.1  
30:00 264.2 F 23.2  
32:00 271.8 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.1  
1:00 272.2 F 28.8  
2:00 271.9 F 28.2  
3:00 271.9 F 28.6  
4:00 272.1 F 28.5  
  
Min 271.3 F 27.9  
Max 272.5 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.1 F 0.0  
5:00 214.0 F 0.0  
10:00 214.6 F 0.0  
15:00 209.2 F 0.0  
20:00 206.9 F 0.0  
25:00 209.8 F 0.0  
30:00 210.9 F 0.0  
Drying Complete

FILLING: 1:11  
HEATING: 32:01  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:10:43  
CYCLE COMPLETE

11/21/2023 10:34 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_